


# SP06PC - Pre-Instructional Survey

Name \_\_\_\_\_ Date \_\_\_\_\_

1. What valve size is the most common in a facility?
  - a. 2 inch
  - b. <1 inch
  - c. >6 inch
  - d. 1 inch
  
2. The monitoring instrument shall be equipped with a probe not to exceed \_\_\_\_\_ in outside diameter?
  - a. 1/2"
  - b. 1/4"
  - c. 1"
  - d. Can be as large as needed
  
3. On a typical P&ID this symbol indicates which of the following 
  - a. check valve
  - b. gate valve
  - c. control valve
  - d. pump
  
4. A monitoring instrument does not have to be calibrated if you're only checking one component.
  - a. True
  - b. False
  
5. The more rings of packing in a valve give you a better chance of stopping a leak?
  - a. True
  - b. False

6. When does a precision calibration have to be done on an instrument?
  - a. Once every quarter
  - b. Prior to place the instrument into service
  - c. When repairs have been made that would change the response time
  - d. All of the above
  
7. Response time is the time between removing the zero gas and to when \_\_\_\_\_ of the calibration gas is attained?
  - a. 95%
  - b. 80%
  - c. 90%
  - d. 100%
  
8. Valve stem finish on a valve is usually checked during the repacking of a valve?
  - a. True
  - b. False
  
9. When being audited by a state or federal agency it is important to answer every question without hesitation?
  - a. True
  - b. False
  
10. Putting two valves back to back (or piggy back) will assure that you will not have an open end line?
  - a. True
  - b. False

## SP06PC - Pre-Instructional Survey Answer Sheet

1. b
2. b
3. a
4. b
5. b
6. d
7. c
8. b
9. b
10. b